

BLANK PAGE



AMENDMENT NO. 1 JANUARY 1993 TO IS 10539: 1983 SPECIFICATION FOR SICKLE TERMINAL DEVICE FOR ARTIFICIAL LIMBS

(Page 1, Fig. 1):

- a) (Height of the sickle) Substitute '75 to 80' for '72 \pm 2'.
- b) (Length of the thread) Substitute '15 to 20' for '20 max'.

(MHD 10)

Reprography Unit, BIS, New Delhi, India



Indian Standard

SPECIFICATION FOR SICKLE TERMINAL DEVICE FOR ARTIFICIAL LIMBS

- 1. Scope Specifies dimensional and other requirements for sickle terminal device for artificial limbs.
- 2. Shape and Dimensions As shown in Fig. 1.




FIG. 1 SICKLE, TERMINAL DEVICE, TYPICAL

- 2.1 The device may be made in one piece or as composite construction in three pieces.
- 2.2 Shape and size of the sickle may vary as agreed to between the purchaser and the manufacturer except that of the interchangeable dimensions of screw thread.
- 3. Material High carbon steel conforming to Designation 30 C8 or 45 C8 of IS: 1875-1978 'Carbon steel billets, blooms, slabs and bars for forgings (fourth revision)'.

4. Workmanship and Finish

- 4.1 The device shall be forged to the given shape.
- 4.2 All surfaces shall be finished smooth and shall be free from sharp edges, nicks and projections, except the cutting edge.
- 4.3 The teeth of the cutting edge shall be sharp, clean and free from burrs, projections, etc.
- 4.4 When made as composite construction, welded or brazed joints shall be dressed smooth.
- 4.5 The complete assembly after finishing shall be given phosphated finish.
- 5. Marking The device shall be marked with the manufacturers name, initials or recognized trade-mark.
- 5,1 ISI Certification Marking Details available with the Indian Standards Institution.
- **6. Packing** The device shall be packed in polyethylene bag individually. The overall packing for multiple numbers may be as agreed to between the purchaser and the manufacturer.

Adopted 18 March 1983 C August 1983, ISI Gr 1